# **PENTYL AMINE**



## 1. PERFORMANCE

1) Measuring range 2-22 ppmNumber of pump strokes  $1(100 \text{m} \ell)$ 

2) Sampling time : 1 minute/1 pump stroke

3) Detectable limit : -4) Shelf life : 3 years 5) Operating temperature :  $15 \sim 25 \,^{\circ}$ C

6) Reading : Graduations printed on the tube are calibrated by Ammonia at 1 pump stroke

and Pentyl amine concentration is determined by using a conversion chart at 1

pump stroke.

7) Colour change : Pale purple → Pale Yellow

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

#### 3. CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured. CH<sub>3</sub> (CH<sub>2</sub>)<sub>3</sub>CH<sub>2</sub>NH<sub>2</sub> + H<sub>3</sub>PO<sub>4</sub> → (R<sub>1</sub>NH<sub>3</sub>)<sub>3</sub>PO<sub>4</sub>

#### 4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

### 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.

